



Shannon Free Zone, Shannon, Co. Clare, Ireland

Tel: +353 61 472622 Fax: +353 61 472642

Email: sales@reagecon.ie

www.reagecon.com

CERTIFICATE OF ANALYSIS

PRODUCT: 10 Platinum Cobalt Colour Standard (Hazen)

PRODUCT NO: HAZ10

LOT NO: HAZ1023C1

DATE OF TEST: 27/03/2023

EXPIRY DATE: 28/03/2025

MEAN ASSAY: 10.0 ± 0.1

SPECIFICATIONS: 9.9 - 10.1 Hazen units

TEST METHOD:

The result reported above was determined by analysis of a sample of this lot taken at time of manufacture.

The spectral characteristics of the sample were recorded in a 10 mm cell between 380nm and 780nm using a Shimadzu UV1800 in accordance with LAB-P-TPUVT. The chromaticity coordinates reported above were calculated from this data according to ASTM E308, using Illuminant C and 2° Observer.

The Pt-Co value of this liquid reference was calculated from the information provided in ASTM D1209 and ISO 6271-2. Analysis of this product is traceable to NIST SRM 930e.

The uncertainty reported is the expanded uncertainty (k=2) for a 95% confidence interval.

This certificate is electronically generated and does not require a signature.

DATE: 29/03/2023

This certificate must not be reproduced except in full.