



Shannon Free Zone,
Shannon, Co. Clare, Ireland
Tel: +353 61 472622
Email: sales@reagecon.ie
Web: www.reagecon.com

CERTIFICATE OF ANALYSIS

PRODUCT: VISCOSITY STANDARD: N14
PRODUCT NO: REVIS-N14
LOT NO: VISN1422J1
DATE OF TEST: 07/11/2022
EXPIRY DATE: 28/09/2025

CERTIFIED VALUES:

Temperature		Viscosity		Density
°C	°F	Kinematic mm ² /s (cSt)	Dynamic mPa.s (cP)	g/ml
20.00	68.00	30.36	25.27	0.8325
25.00	77.00	24.66	20.45	0.8293
37.78	100.00	15.40	12.64	0.8210
40.00	104.00	14.30	11.72	0.8196
50.00	122.00	10.52	8.555	0.8131

Method of Analysis: Kinematic Viscosity measurements are in accordance with ASTM D2162.
Density measurements are in accordance with ASTM D4052.
(Dynamic Viscosity = Kinematic Viscosity x Density)

Metrological Traceability: Kinematic Viscosity and Density values are traceable to the Primary Standard of pure water at 20°C, as adopted by NIST. Temperature values are traceable via National Standards to the ITS-90 temperature scale.

EXPANDED UNCERTAINTIES ($k=2$)

Viscosity Range (mm ² /s)	Kinematic Viscosity (mm ² /s)	Dynamic Viscosity (mPa.s)
0.25 - 6.6	0.14%	0.22%
6.6 - 14	0.16%	0.23%
14 - 37	0.17%	0.24%
37 - 85	0.19%	0.25%
85 - 230	0.21%	0.27%
230 - 540	0.23%	0.29%

Viscosity Range (mm ² /s)	Kinematic Viscosity (mm ² /s)	Dynamic Viscosity (mPa.s)
540 - 1500	0.24%	0.29%
1500 - 3600	0.26%	0.31%
3600 - 9000	0.27%	0.32%
9000 - 25500	0.30%	0.34%
25500 - 110000	0.31%	0.35%

The reported analysis was performed by a third party on behalf of Reagecon. This certificate and its reported results are not covered by Reagecon's scope of accreditation to ISO 17025 (INAB Ref. 264T)

Date: 08/11/2022 This certificate is generated electronically and does not require a signature
This certificate must not be reproduced except in full.