



Certificate of Analysis

Turbidimetric Calibration Standard

4 NTU Ratio Turbidity Standard

Product No:	CRSR-4-500
Lot No:	CRSR420H1
Expiry Date:	28/08/2022

Certified Value:	3.9 NTU
Date of Measurement:	10/08/2020

Assay Limits:
3.8 - 4.2 NTU

Method:
Tested by Turbidity Measurement to In-House Method TPTURB.

Traceability:
The certified value is directly traceable to the Formazin Primary Standard for Turbidity, as adopted by NIST and other National Metrology Institutes.

Reference:
Water Quality-Determination of Turbidity - International Standard ISO 17027, DIN 38 404 and NF EN 27027.

Products are manufactured under an NSAI registered I.S EN ISO9001:2015 Quality System, registration no: 19.2769

Date of Issue of the Certificate: 10/08/2020

Eileen Malone

QC Technician Eileen Malone

Complementary information relative to this product is available at www.reagecon.com
Page 1 of 1. This Certificate must not be reproduced except in full. Rev-20H1

Reagecon Diagnostics Ltd.
Shannon Free Zone, Shannon, Co. Clare, Ireland.
Tel +353 61 472622, Fax: +353 61 472642
Email: sales@reagecon.ie, www.reagecon.com