

ENVIRONMENTAL MATRIX REFERENCE MATERIAL

TMDA-51.5, lot 0621

Product Information Sheet

A Trace Elements Fortified Sample



Updated: 2021-JUN-15

Trace element matrix Reference Materials (RMs) are made with filtered and diluted Lake Ontario water and are preserved with 0.2% nitric acid. This fortified bulk RM has concentrations in the high range.

This Reference Material is intended for the verification or development of analytical methods for environmental analysis. It is not intended for use as a calibration standard.

Measurand	Assigned Value in µg/L*		C.I.	N
Aluminum	105	±	9	151
Antimony	15.2	±	1.3	131
Arsenic	16.7	±	1.4	154
Barium	72.6	±	4.1	149
Beryllium	9.99	±	0.92	142
Bismuth	9.79	±	2.13	70
Boron	46.9	±	5.6	98
Cadmium	25.7	±	1.7	161
Chromium	65.1	±	4.3	160
Cobalt	68.7	±	4.7	148
Copper	77.4	±	5.4	162
Gallium	9.83	±	0.70	47
Iron	108	±	10	143
Lead	63.7	±	5.6	155
Lithium	19.4	±	1.9	98
Manganese	78.7	±	4.4	148
Molybdenum	55.0	±	4.5	139
Nickel	65.3	±	4.3	154
Rubidium	15.3	±	1.0	62
Selenium	14.3	±	1.9	134
Silver	12.1	±	1.1	129
Strontium	118	±	7	136
Thallium	20.3	±	1.8	128
Tin	16.4	±	1.4	113
Titanium	14.1	±	1.0	101
Uranium	28.7	±	2.3	120
Vanadium	46.5	±	3.5	142

*Unless otherwise indicated

Assigned values are the robust means calculated using Algorithm A (ISO/IEC 13528). Confidence interval (C.I.) represents +/- 2 times the calculated robust standard deviation (R.STD) of all reported laboratory values, calculated by Algorithm A (ISO/IEC 13528).



Traceability

The stated values are derived from analysis of the inter-laboratory consensus data from the accredited Environment and Climate Change Canada Proficiency Testing studies. (A2LA scope 2867.01)

Methods

Report details of measurement methods used for specific parameters are available upon request.

Storage and Handling

RMs should be stored refrigerated, well sealed and in the dark. Care should be taken when subsampling to avoid contamination of the sample bottle. An aliquot should be removed as necessary and any remaining aliquot sample should be put to waste as appropriate. We strongly recommend that the RM be tightly capped and refrigerated immediately after use.

Expiry Date

Please note that expiry dates of 2 year from the date of shipping are indicative of sample stability, sample transport, handling, and storage. Expiry dates are indicated directly on the sample bottle label. Environment and Climate Change Canada is not liable for samples beyond the stated expiry on the bottle label.

Disclaimer, Liability & Warranty

Environment and Climate Change Canada warrants that the materials conform to the stated values for the duration of the sample validity period. In the event of a breach of this warranty, Environment and Climate Change Canada will only be liable for a replacement sample, an equivalent substitute, or the invoice price of the RM during the period of sample validity. In no event will Environment and Climate Change Canada be liable for direct, indirect, special, incidental or consequential damages arising from the use of or inability to use the material or documentation, or for the loss of revenue or profit, even if advised of the possibility of such damages. Environment and Climate Change Canada's liability does not extend to third party purchasers.

Further Information

Additional information is available on request. Analytical results, any comments or suggestions are welcome. Difficulties or discrepancies arising with the reference materials should be communicated immediately.

Quality Manager

Sarah Pomfret
ECCC RM Programs Signing Officer
E-mail: sarah.pomfret@canada.ca

Customer Service

CRM Sales
Information and Quality Management Group
Environmental Science & Technology
Laboratories Division
Water Science & Technology Directorate
Environment and Climate Change Canada
867 Lakeshore Road
Burlington, ON
L7S 1A1
CANADA
Tel: +1 905-336-4653
Fax: +1 905-336-8914
E-mail: ec.CRMsales.ec@canada.ca

