

Certificate of Analysis

Turbidimetric Calibration Standard

1000 NTU Ratio Turbidity Standard

Product No: CRSR-1000-100
Lot No: CRSR1M21C1
Expiry Date: 28/03/2023

Certified Value: 1020 NTU
Date of Measurement: 18/03/2021
Assay Limits:
950 - 1050 NTU

Method:

Tested by Turbidity Measurement to In-House Method TPTURB.

Traceability:

The certified value is directly traceable to the Formazin Primary Standard for Turbidity, as adopted by NIST and other National Metrology Institutes.

Reference:

Water Quality-Determination of Turbidity - International Standard ISO 17027, DIN 38 404 and NF EN 27027.

Products are manufactured under an NSAI registered I.S EN ISO9001:2015 Quality System, registration no: 19.2769

Date of Issue of the Certificate: 18/03/2021



QC Technician Orla Sheehan

Complementary information relative to this product is available at www.reagecon.com
Page 1 of 1. This Certificate must not be reproduced except in full. Rev-21C40

Reagecon Diagnostics Ltd.
Shannon Free Zone, Shannon, Co. Clare, Ireland.
Tel +353 61 472622, Fax: +353 61 472642
Email: sales@reagecon.ie, www.reagecon.com