

Certificate of Analysis

Conductivity Standard Solution

147 μ S/cm @25°C

Product No:
Lot No:
*Expiry date * :*

**CSKCS
CSKCS22B1
28/02/2023**

Mean specific conductance:
Date of measurement
Date of Sample Receipt:
*Date of Manufacturing * :*

**147.5 μ S/cm @25°C
17/02/2022
17/02/2022
17/02/2022**

* The detail above is based on information supplied in writing by Reagecon Manufacturing.

Specification:

145.5 - 148.5 μ S/cm @25°C

Method:

This certificate relates solely to the sample as received by the laboratory, bearing the product code and lot number given above, taken at time of manufacture. Test Method used was TPCOND. Measurement taken by comparison with standard prepared from National Institute of Standards and Technology (USA), Standard Reference Material 999 (Potassium Chloride). Electrode used for measurement: Platinised Platinum Dip Cell. Reference: ASTM D-1125 Method A. The uncertainty of measurement has been calculated not to exceed $\pm 1\%$ at 95% confidence level, $k=2$.

Accreditation:

Reagecon Diagnostics Ltd. is accredited by the Irish National Accreditation Board, under scope 264T, for the test method, TPCOND, used to generate the above result. This accreditation is intended only to certify that Reagecon has the Quality Management Systems in place to ensure that each individual test result generated using TPCOND is technically valid and is supported by appropriate uncertainty measurements.

Date of issue of the certificate

17/02/2022



QC Technician Patricia de Guzman

Complementary information relative to this product is available at www.reagecon.com
Page 1 of 1. This Certificate must not be reproduced except in full. Rev-22B15

Tested by Reagecon Quality Control
Laboratories for Reagecon Manufacturing

Reagecon Diagnostics Ltd.
Shannon Free Zone, Shannon, Co. Clare, Ireland.
Tel +353 61 472622, Fax: +353 61 472642
Email: sales@reagecon.ie, www.reagecon.com