



Certificate of Analysis

Turbidimetric Calibration Standard

5 NTU Ratio Turbidity Standard

Product No:	CRSR-5-500
Lot No:	CRSR521L1
Expiry date * :	28/11/2023

Certified Value:	5.01 NTU
Date of Measurement:	01/12/2021
Date of Sample Receipt:	01/12/2021
Date of Manufacturing * :	01/12/2021

* The detail above is based on information supplied in writing by Reagecon Manufacturing.

Assay Limits:

4.75 - 5.25NTU

Method:

Tested by Turbidity Measurement to In-House Method TPTURB.

Traceability:

The certified value is directly traceable to the Formazin Primary Standard for Turbidity, as adopted by NIST and other National Metrology Institutes.

Reference:

Water Quality-Determination of Turbidity - International Standard ISO 17027, DIN 38 404 and NF EN 27027.

Date of Issue of the Certificate: 01/12/2021

A handwritten signature in blue ink, appearing to read 'Orla Sheehan'.

QC Technician **Orla Sheehan**

Complementary information relative to this product is available at www.reagecon.com
Page 1 of 1. This Certificate must not be reproduced except in full. Rev-21L25

Tested by Reagecon Quality Control
Laboratories for Reagecon Manufacturing

Reagecon Diagnostics Ltd.
Shannon Free Zone, Shannon, Co. Clare, Ireland.
Tel +353 61 472622, Fax: +353 61 472642
Email: sales@reagecon.ie, www.reagecon.com